

*Please note – For all information regarding modifications and deviations from ISO22196:2011, please request statement B-Lab-State-007

| Noted during testing | | |
|--------------------------------------|---|-----------|
| | Note | Deviation |
| No. of samples supplied per test | 6 of 6 required | NO |
| Sample size | Average size: 5x5cm <i>Required size: 5x5cm</i> | N/A |
| Sample sterilisation method | DRY HEAT ONLY | N/A |
| Film size | Average size: 4x4cm <i>Required size: 4x4cm</i> | N/A |
| Inoculation volume | Volume: 0.1ml <i>Volume required by method: 0.1-0.4ml</i> | NO |
| Inoculation concentration at initial | Concentration of cells/ml: MRSA: 6.22×10^5 <i>E. coli: 8.51×10^5</i> <i>P. aeruginosa: 4.15×10^5</i> <i>Concentration required by method:</i> <i>$2.5 \times 10^5 - 10 \times 10^5$ cells/ml</i> | NO |



Megan Vaughan (BSc)
Laboratory Manager,
BioLabTests Ltd

